

# United States Food and Drug Administration (FDA) Medical Device Classification System



**Stuart R. Goldman**  
United States Senior Consultant, RA/QA  
ASQ (CMQ/OE, CMDA, CQA)  
stuart.goldman@ul.com

June 2023



by UL


# What you will learn in this white paper

- History of medical device classification in the U.S.
- FDA medical device classification system
- Classifying your medical device and some helpful tips in the process

## Medical device classification in the U.S.

In 1976, the Medical Device Amendments were added to the United States (U.S.) Federal Food, Drug and Cosmetic Act (FD&C Act) of 1938, establishing the current risk-based classification system for all medical devices the FDA regulates. The Medical Device Amendments were intended to provide reasonable assurance of the safety and effectiveness of all medical devices.

The Medical Device Amendments also established that, for devices that were not on the market prior to passage of the amendments on May 28, 1976 — aka pre-amendment devices — or for devices that had been significantly modified, a premarket notification (510(k)), De Novo or premarket approval (PMA) pathway would be required to place these higher-risk devices on the market.



# The FDA medical device classification system

Knowing the correct classification of your **medical device** is crucial for mapping its regulatory route to commercialization in the U.S. How the FDA classifies a medical device depends on the level of risk and **Regulatory Controls** appropriate to control the risk the device poses to the intended patient and user(s) – e.g., lay users, healthcare professionals, and caregivers. The risks and regulatory controls applicable to a medical device are specific to its intended use, indications for use and technological

characteristics as these things can influence the time and costs of bringing your device to market. Higher-risk devices are significantly more costly and take longer to bring to market than lower-risk devices.

There are three main classes of medical devices in the U.S. (Class I, Class II, and Class III) in addition to “unclassified” (aka pre-amendment devices), “not classified” and “humanitarian device exemption” (HDE) devices. Every medical device type has a unique three-letter

product code and name attached to it. The FDA<sup>1</sup> currently has approximately 6,500 product codes assigned that reside within the different **Device Classification Panels** found in Parts 862-892 of the Code of Federal Regulations (CFR), for which the FDA has also established more than 1,700 seven-digit regulation numbers to describe the different device types. The number of product codes increases every year.





**Class I** devices are low-risk, only require general controls and typically do not require 510(k) clearance from the FDA under a premarket notification, as most of these devices are 510(k)-exempt. However, there are some Class I devices that do require 510(k) clearance.

**Class II** devices are moderate-risk and require general and special controls and 510(k) clearance from the FDA. However, there are some Class II devices that do not require 510(k) clearance.

Note that for those Class I and Class II devices that the FDA has exempted from the 510(k) requirements as described on the product classification page for the related product code, if the device exceeds the limitations of device exemptions covered under 21 CFR XXX.9, where XXX represents the review panel number — e.g., 872 Dental — a 510(k) may be required and the specific parts of XXX.9 (a, b, c) should be reviewed to confirm their relevance. As described in XXX.9, (a) is for a different intended use; (b) is for a different fundamental scientific technology; and (c) is for in vitro diagnostic (IVD) devices, with certain stipulations as described in the regulation.

**Class III** devices entail the highest risk and require general controls and approval from the FDA under a PMA application. These devices require clinical data to provide the FDA with adequate assurance of safety and effectiveness.

For devices that are not high-risk but for which no legally marketed predicate device exists, a De Novo submission can be made to the FDA. Under a De Novo pathway, the FDA will reclassify the device as either Class I or Class II, for which general controls of the FD&C Act apply, as well as device-specific special controls agreed to by the FDA and the device sponsor for Class II devices. Once a De Novo has been granted for the new device and assigned a DEN number, e.g., DEN123456, the FDA assigns a new product code and may assign a new regulation number into the federal register, allowing the device to be sold on the market. The new device can then be used as a predicate device for subsequent submissions under a 510(k).

It is important to remember that, while each medical device has a three-letter product code assigned to it and there are more than 6,500 product codes in the FDA product classification database, there are approximately 1,700<sup>1</sup> regulation numbers. Therefore, there are many examples where multiple product codes fall within the same regulation number. In some cases, there are more than 50 product codes assigned to the same regulation number. This makes selecting the correct product code challenging, as further discussed in this white paper. In addition, there are many devices for which no regulation has been established.

**Figures 1-8** show the product classification pages for selected Class I, Class II and Class III devices, as well as for unclassified, not classified and HDE devices, to give the reader some basic familiarity with the type of regulatory information that the FDA includes on these pages, which differs depending on the device's classification, as seen in the following examples.

Figure 1 – Class I Device Product Classification Page — Product Code FOB (Bedpan)

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[New Search](#)
[Back to Search Results](#)

|   |  |
|---|--|
| <b>Device</b>   | Bedpan   |
| <b>Regulation Medical Specialty</b>   | General Hospital   |
| <b>Review Panel</b>   | General Hospital   |
| <b>Product Code</b>   | FOB  |
| <b>Premarket Review</b>   | <a href="#">Drug Delivery and General Hospital Devices, and Human Factors (DHT3C)</a><br>Drug Delivery and General Hospital Devices, and Human Factors (DHT3C) |
| <b>Submission Type</b>  | 510(K) Exempt  |
| <b>Regulation Number</b>  | 880.6730   |
| <b>Device Class</b>   | 1  |
| <b>Total Product Life Cycle (TPLC)</b>  | <a href="#">TPLC Product Code Report</a>   |
| <b>GMP Exempt?</b>  | Yes  |
| <p><b>Note:</b> This device is also exempted from the GMP regulation, except for general requirements concerning records (820.180) and complaint files (820.198), <i>as long as the device is <u>not</u> labeled or otherwise represented as sterile.</i></p>   |  |
| <b>Summary Malfunction Reporting</b>  | Eligible   |
| <p><b>Note:</b> FDA has exempted almost all class I devices (with the exception of <a href="#">reserved devices</a>) from the premarket notification requirement, including those devices that were exempted by final regulation published in the <i>Federal Registers</i> of December 7, 1994, and January 16, 1996. It is important to confirm the exempt status and any limitations that apply with <a href="#">21 CFR Parts 862-892</a>. Limitations of device exemptions are covered under 21 CFR XXX.9, where XXX refers to Parts 862-892.</p> <p>If a manufacturer's device falls into a generic category of exempted class I devices as defined in <a href="#">21 CFR Parts 862-892</a>, a premarket notification application and fda clearance is not required before marketing the device in the U.S. however, these manufacturers are required to register their establishment. Please see the <a href="#">Device Registration and Listing website</a> for additional information.</p> |  |
| <b>Implanted Device?</b>  | No   |
| <b>Life-Sustain/Support Device?</b>   | No   |
| <b>Third Party Review</b>   | Not Third Party Eligible   |

Page Last Updated: 05/22/2023

Figure 2 – Class II Device Product Classification Page — Product Code FRG (Wrap, Sterilization)

## Product Classification

[FDA Home](#) [Medical Devices](#) [Databases](#)

[New Search](#) [Back to Search Results](#)

|                                 |  |
|---------------------------------|--|
| Device                          | Wrap, Sterilization  |
| Regulation Medical Specialty    | General Hospital   |
| Review Panel                    | Orthopedic   |
| Product Code                    | FRG  |
| Premarket Review                | <a href="#">Spinal Devices (DHT6B)</a><br><a href="#">Spinal Devices (DHT6B)</a>   |
| Submission Type                 | 510(k)   |
| Regulation Number               | 880.6850   |
| Device Class                    | 2  |
| Total Product Life Cycle (TPLC) | <a href="#">TPLC Product Code Report</a>   |
| GMP Exempt?                     | No   |
| Summary Malfunction Reporting   | Eligible   |
| Implanted Device?               | No   |
| Life-Sustain/Support Device?    | No   |
| Third Party Review              | <ul style="list-style-type: none"><li>• <a href="#">Eligible for 510(k) Third Party Review Program</a></li></ul>   |
| Accredited Persons              | <ul style="list-style-type: none"><li>• <a href="#">Aabb</a></li><li>• <a href="#">Center For Measurement Standards Of Industrial</a></li><li>• <a href="#">Global Quality And Regulatory Services</a></li><li>• <a href="#">Regulatory Technology Services, Llc</a></li><li>• <a href="#">Third Party Review Group, Llc</a></li></ul> |

Page Last Updated: 05/22/2023

Figure 3 – Class III Device Product Classification Page — Product Code LWQ (Heart-Valve, Mechanical)

## Product Classification

[FDA Home](#) [Medical Devices](#) [Databases](#)

[New Search](#) [Back to Search Results](#)

|                                 |   |
|---------------------------------|---|
| Device                          | Heart-Valve, Mechanical   |
| Review Panel                    | Cardiovascular  |
| Product Code                    | LWQ   |
| Premarket Review                | <a href="#">Circulatory Support, Structural and Vascular Devices (DHT2B)</a><br>Circulatory Support, Structural and Vascular Devices (DHT2B)  |
| Submission Type                 | PMA   |
| Device Class                    | 3   |
| Total Product Life Cycle (TPLC) | <a href="#">TPLC Product Code Report</a>  |
| GMP Exempt?                     | No  |
| Summary Malfunction Reporting   | Eligible  |
| Implanted Device?               | Yes   |
| Life-Sustain/Support Device?    | Yes   |
| Recognized Consensus Standards  | <ul style="list-style-type: none"><li>3-174 ISO 5840-1 Second edition 2021-01<br/><a href="#">Cardiovascular implants - Cardiac valve prostheses - Part 1: General requirements</a></li><li>3-175 ISO 5840-2 Second edition 2021-01<br/><a href="#">Cardiovascular implants - Cardiac valve prostheses - Part 2: Surgically implanted heart valve substitutes</a></li></ul> |
| Third Party Review              | Not Third Party Eligible  |

Page Last Updated: 05/22/2023

Unclassified or pre-amendment devices are devices that existed before the establishment of the medical device regulations on May 28, 1976. There are 85<sup>1</sup> product codes for these devices, and none of them has a class (Class I, II or III) or regulation number (CFR 123.4567) assigned, as seen in **Figure 4**, using product code FRO as an example. Most unclassified devices require 510(k) clearance, while some are under enforcement discretion as defined on their product classification page for submission type. For product code FRO, the submission type is a 510(k).

**Figure 4 – Unclassified Device Product Classification Page — Product Code FRO (Dressing, Wound, Drug)**

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[New Search](#)
[Back to Search Results](#)

|  |  |
|--|--|
| <b>Device</b>                          | Dressing, Wound, Drug  |
| <b>Review Panel</b>                    | General & Plastic Surgery  |
| <b>Product Code</b>                    | FRO  |
| <b>Premarket Review</b>                | <a href="#">Infection Control and Plastic Surgery Devices (DHT4B)</a><br><a href="#">Infection Control and Plastic Surgery Devices (DHT4B)</a> |
| <b>Unclassified Reason</b>             | Pre-Amendment  |
| <b>Submission Type</b>                 | 510(k)   |
| <b>Device Class</b>                    | Unclassified   |
| <b>Total Product Life Cycle (TPLC)</b> | <a href="#">TPLC Product Code Report</a>   |
| <b>GMP Exempt?</b>                     | No   |
| <b>Summary Malfunction Reporting</b>   | Ineligible   |
| <b>Implanted Device?</b>               | No   |
| <b>Life-Sustain/Support Device?</b>    | No   |
| <b>Third Party Review</b>              | Not Third Party Eligible   |

Page Last Updated: 05/22/2023



Not-classified devices are post-amendment devices, and many of them are under emergency use authorization (EUA), enforcement discretion (ED) or for export only (FEO). There are currently 369<sup>1</sup> product codes for these devices, and none of them has a class or regulation number assigned, as seen in **Figures 5-7** for product codes QLW, PFU and PRF, respectively.

**Figure 5 – Not Classified Device Product Classification Page — Product Code QLW (Covid-19 Test Home Collection Kit Devices), Under Emergency Use Authorization**

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[New Search](#)
[Back to Search Results](#)

|  |   |
|--|---|
| <b>Device</b>                          | Covid-19 Test Home Collection Kit Devices   |
| <b>Definition</b>                      | Specimens collected using the Home Collection Kit can be transported at ambient temperature for testing at a laboratory. SARS-CoV-2 RNA from the clinical specimen is maintained in the specimen packaging and suitable for use in diagnostic testing performed using a molecular in vitro diagnostic (IVD) test for the detection of SARS-CoV-2 RNA that is authorized for use with the COVID-19 Test Home Collection Kit. |
| <b>Physical State</b>                  | Specimen collection device, swab and/or tube, stabilizing reagents.   |
| <b>Technical Method</b>                | Collection and maintenance of nucleic acid from SARS-CoV-2 in clinical specimens  |
| <b>Target Area</b>                     | The device is a specimen collection device; none of the body parts will utilize the device or are intended to be affected by the device.  |
| <b>Review Panel</b>                    | Microbiology  |
| <b>Product Code</b>                    | QLW   |
| <b>Premarket Review</b>                | <a href="#">Division of Microbiology Devices (DMD)</a><br>Division of Microbiology Devices (DMD)  |
| <b>Not Classified Reason</b>           | EUA - Emergency Use Authorization   |
| <b>Submission Type</b>                 | EUA - Emergency Use Authorization   |
| <b>Device Class</b>                    | Not Classified  |
| <b>Total Product Life Cycle (TPLC)</b> | <a href="#">TPLC Product Code Report</a>  |
| <b>GMP Exempt?</b>                     | No  |
| <b>Summary Malfunction Reporting</b>   | Ineligible  |
| <b>Implanted Device?</b>               | No  |
| <b>Life-Sustain/Support Device?</b>    | No  |
| <b>Third Party Review</b>              | Not Third Party Eligible  |

Page Last Updated: 05/22/2023

Figure 6 – Not Classified Device Product Classification Page — Product Code PFU (Tissue Processing Kit), Under Enforcement Discretion

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[New Search](#)
[Back to Search Results](#)

|  |   |
|--|---|
| <b>Device</b>                          | Tissue Processing Kit   |
| <b>Definition</b>                      | For the preparation of human tissue specimens for frozen sectioning and histological examination.   |
| <b>Physical State</b>                  | Freezing device, heat sink and a chuck with embeddign mold  |
| <b>Technical Method</b>                | The freezing device and heat sink are housed in an insulated box and the entire kit would be used at the dissecting bench, allowing users to freeze the sample for future cutting or slicing. |
| <b>Target Area</b>                     | Any human tissue  |
| <b>Review Panel</b>                    | Pathology   |
| <b>Product Code</b>                    | PFU   |
| <b>Premarket Review</b>                | <a href="#">Division of Molecular Genetics and Pathology (DMGP)</a><br>Division of Molecular Genetics and Pathology (DMGP)  |
| <b>Not Classified Reason</b>           | Enforcement Discretion  |
| <b>Submission Type</b>                 | Enforcement Discretion  |
| <b>Device Class</b>                    | Not Classified  |
| <b>Total Product Life Cycle (TPLC)</b> | <a href="#">TPLC Product Code Report</a>  |
| <b>GMP Exempt?</b>                     | No  |
| <b>Summary Malfunction Reporting</b>   | Eligible  |
| <b>Implanted Device?</b>               | No  |
| <b>Life-Sustain/Support Device?</b>    | No  |
| <b>Third Party Review</b>              | Not Third Party Eligible  |

Page Last Updated: 05/22/2023

Figure 7 – Not Classified Device Product Classification Page — Product Code PZN (Placental Growth Factor Test), For Export Only

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

New Search
Back to Search Results

|  |  |
|--|--|
| <b>Device</b>                          | Placental Growth Factor Test   |
| <b>Definition</b>                      | For quantitative determination of Placental Growth Factor (PIGF).  |
| <b>Physical State</b>                  | In Vitro Diagnostic (IVD)  |
| <b>Technical Method</b>                | Immunoassay  |
| <b>Target Area</b>                     | In Vitro Diagnostic (IVD)  |
| <b>Review Panel</b>                    | Toxicology   |
| <b>Product Code</b>                    | PZN  |
| <b>Premarket Review</b>                | <a href="#">Division of Chemistry and Toxicology Devices (DCTD)</a><br><a href="#">Division of Chemistry and Toxicology Devices (DCTD)</a> |
| <b>Not Classified Reason</b>           | For Export Only  |
| <b>Submission Type</b>                 | Contact ODE  |
| <b>Device Class</b>                    | Not Classified   |
| <b>Total Product Life Cycle (TPLC)</b> | <a href="#">TPLC Product Code Report</a>   |
| <b>GMP Exempt?</b>                     | No   |
| <b>Summary Malfunction Reporting</b>   | Ineligible   |
| <b>Implanted Device?</b>               | No   |
| <b>Life-Sustain/Support Device?</b>    | No   |
| <b>Third Party Review</b>              | Not Third Party Eligible   |

Page Last Updated: 05/22/2023

A humanitarian device exemption (HDE) is a marketing application for a humanitarian use device (HUD). A HUD is a medical device intended to benefit patients undergoing treatment or diagnosis of a disease or condition that affects or is manifested in not more than 8,000 individuals in the U.S. per year. There are currently 48<sup>1</sup> product codes for these devices, as seen in **Figure 8** for product code MWH.

**Figure 8 – HDE Device Product Classification Page — Product Code MWH (Pulmonic Valved Conduit)**

**Product Classification**  
FDA Home Medical Devices Databases

[New Search](#) [Back to Search Results](#)

|                                 |  |
|---------------------------------|--|
| Device                          | Pulmonic Valved Conduit  |
| Review Panel                    | Cardiovascular   |
| Product Code                    | MWH  |
| Premarket Review                | <a href="#">Circulatory Support, Structural and Vascular Devices (DHT2B)</a><br>Circulatory Support, Structural and Vascular Devices (DHT2B) |
| Submission Type                 | HDE - Humanitarian Device Exemption  |
| Device Class                    | HDE  |
| Total Product Life Cycle (TPLC) | <a href="#">TPLC Product Code Report</a>   |
| GMP Exempt?                     | No   |
| Summary Malfunction Reporting   | Eligible   |
| Implanted Device?               | Yes  |
| Life-Sustain/Support Device?    | Yes  |
| Third Party Review              | Not Third Party Eligible   |

Page Last Updated: 05/22/2023



## Classifying your medical device

With a basic understanding of the various categories the FDA uses to classify medical devices as described above, there are many different ways to classify a medical device using certain FDA databases. Those databases are linked below. There are times when a certain database will prove more helpful than another, and in some cases, using more than one database will help further confirm your device's classification to another legally marketed device in the U.S., based on its intended use and technological characteristics.

- The [Product Classification](#) database
- The [Registration & Listing](#) database
- The [510\(k\)](#) database
- The [De Novo](#) database
- The [PMA](#) database
- The [HDE](#) database

## The FDA Product Classification database

The FDA Product Classification database is best used if you are unsure how your device should be classified based on its intended use and technological characteristics. When using this database, you should select keywords that are specific to your device with respect to its intended use and technology as an initial starting point to see what product codes and regulation numbers are extracted from the database.

There are two options for using the Product Classification database as seen in **Figure 9**: the Quick Search option (top) and the Advanced Search option (bottom). Using the advanced search option allows for more customization in your search as you can use other drop-down menus — e.g., review panel, implanted device, etc., as seen below — besides using just the Device field. This can help focus your search efforts beyond what you are limited to when using the Quick Search option.



It is important to note that using the same word(s) in the Quick Search and Advanced Search options will often produce a different number of product codes, as seen in **Table 1** and **Figure 10** when using “catheter,” “condom,” “shoulder prosthesis,” “infusion pump” and “artificial intelligence” as some examples.

**Table 1 – Product Classification Database Comparison — Quick Search Option vs. Advanced Search Option**

| Search word(s)          | Quick search results | Advanced search results |
|-------------------------|----------------------|-------------------------|
| Catheter                | 269                  | 262                     |
| Condom                  | 11                   | 12                      |
| Shoulder prosthesis     | 11                   | 14                      |
| Infusion pump           | 27                   | 22                      |
| Artificial intelligence | 9                    | 1                       |

Let us look at three different examples from Table 1.

### Classification Example 1 – Condom devices

The figure below shows the subtle differences when using the Quick Search option and Advanced Search option for the word “condom.” It is also important not to confuse the product code name with the regulation number name, although they can often be the same, as seen below.

**Figure 10 – Product Classification Database Comparison — Quick Search Option (Top) vs. Advanced Search Option (Bottom) for “Condom”**

The screenshot displays the search results for 'Condom' in the Product Classification database. The table has the following columns: Product Code, Device, Regulation Number, and Device Class. Red lines and dots highlight the differences between the 'Device' column (Quick Search results) and the 'Regulation Name' column (Advanced Search results).

| Product Code        | Device  | Regulation Name                   | Regulation Number | Device Class |
|---------------------|---|-----------------------------------|-------------------|--------------|
| <a href="#">HIS</a> | <a href="#">Condom</a>  | Condom                            | 884.5300          | 2            |
| <a href="#">LTZ</a> | <a href="#">Condom With Nonoxynol-9</a>                                     | Condom With Spermicidal Lubricant | 884.5310          | 2            |
| <a href="#">OBY</a> | <a href="#">Condom_Female_Animal Tissue</a>                                 | Multiple-Use Female Condom        | 884.5330          | 3            |
| <a href="#">MOL</a> | <a href="#">Condom_Synthetic</a>  | Condom                            | 884.5300          | 2            |
| <a href="#">QRZ</a> | <a href="#">External Condom For Anal Intercourse Or Vaginal Intercourse</a> |                                   | 884.5305          | 2            |
| <a href="#">QBB</a> | <a href="#">Kit_Conception-Assist_Home Use</a>                              | Cervical Cap                      | 884.5250          | 2            |
| <a href="#">NUC</a> | <a href="#">Lubricant_Personal</a>  | Condom                            | 884.5300          | 2            |
| <a href="#">PEB</a> | <a href="#">Lubricant_Personal_Gamele_Fertilizati...</a>                    | Condom                            | 884.5300          | 2            |
| <a href="#">LZL</a> | <a href="#">Micro-Condom</a>  | Glans Sheath                      | 884.5320          | 3            |
| <a href="#">QPD</a> | <a href="#">Personal Lubricant Ring</a>                                     | Condom                            | 884.5300          | 2            |
| <a href="#">MBU</a> | <a href="#">Single-Use Internal Condom</a>                                  | Single-Use Internal Condom        | 884.5340          | 2            |

Page Last Updated: 05/22/2023

Figure 10 – Product Classification Database Comparison — Quick Search Option (Top) vs. Advanced Search Option (Bottom) for “Condom” – Continued

**Product Classification**

FDA Home Medical Devices Databases

1 to 12 of 12 results  
condom

Results per page 25

New Search Export to Excel Help

| Product Code | Device  | Regulation Number                 | Device Class |   |
|--------------|---|-----------------------------------|--------------|---|
| LZL          | Micro-Condom  | Glans Sheath                      | 884.5320     | 3 |
| OBY          | Condom, Female, Animal Tissue                               | Multiple-Use Female Condom        | 884.5330     | 3 |
| HIS          | Condom  | Condom                            | 884.5300     | 2 |
| LTZ          | Condom With Nonoxynol-9                                     | Condom With Spermicidal Lubricant | 884.5310     | 2 |
| MBU          | Single-Use Internal Condom                                  | Single-Use Internal Condom        | 884.5340     | 2 |
| MOL          | Condom, Synthetic   | Condom                            | 884.5300     | 2 |
| MSC          | Barrier, Std, Oral Sex                                      | Condom                            | 884.5300     | 2 |
| NUC          | Lubricant, Personal   | Condom                            | 884.5300     | 2 |
| OKW          | Seminal Fluid Collection Kit                                | Condom                            | 884.5300     | 2 |
| PEB          | Lubricant, Personal, Gamete, Fertilization, And Em...       | Condom                            | 884.5300     | 2 |
| QPD          | Personal Lubricant Ring                                     | Condom                            | 884.5300     | 2 |
| QRZ          | External Condom For Anal Intercourse Or Vaginal Intercourse |                                   | 884.5305     | 2 |

Page Last Updated: 05/22/2023

In the example above, 11 product codes were extracted from the database using the Quick Search option, and 12 product codes were extracted using the Advanced Search option. Only the Quick Search option retrieved product codes MSC (barrier, STD, oral sex) and OKW (seminal fluid collection kit), while only the advanced search option retrieved product code OBB (kit, conception-assist, home use). All the product codes shown in Figure 10 in this search example fall under Part 884 Obstetrics/Gynecology, covering the following seven regulation numbers:

- [884.5250](#) – Cervical cap
- [884.5300](#) – Condom
- 884.5305 – The regulation name and description are pending with the FDA as product code QRZ, and the regulation number was established under a De Novo pathway (DEN210034).<sup>2</sup>
- [884.5310](#) – Condom with spermicidal lubricant
- [884.5320](#) – Glans sheath
- [884.5330](#) – Multiple-use female condom
- [884.5340](#) – Single-use internal condom

From Example 1, we see that all condoms are not the same, based on their product code name and regulation name. This is where a more detailed review of the intended use and technology of a particular type of condom needs to be undertaken to confirm its correct product code and regulation number. Two of the product codes identified in Figure 10, OBY (condom, female, animal tissue) and LZL (micro-condom), are Class III devices requiring FDA approval under a PMA submission. All the other product codes identified in this example are Class II devices requiring 510(k) clearance. It is also noted that six of the devices listed above do not appear to even be condoms, based on their product code names and other descriptive information found on their respective product classification pages when compared to product code HIS (condom) and its regulation number, 884.5300. Those devices are:

- Product code MSC (barrier, STD, oral sex)/884.5300
- Product code NUC (lubricant, personal)/884.5300
- Product code OBB (kit, conception-assist, home use)/884.5250
- Product code PEB (lubricant, personal, gamete, fertilization, and embryo-compatible)/884.5300
- Product code QPD (personal lubricant ring)/884.5300
- Product code OKW (seminal fluid collection kit)/884.5300



As product codes NUC and QPD, which are personal lubricants in different forms, share the same regulation as condoms in 884.5300, their product classification pages seen in **Figure 11** have additional descriptive information (i.e., description, physical state, technical method and target area) that differentiates them from a condom because the description of a condom found in 884.5300 makes no reference to any type of lubricant. However, since personal lubricants are often used with condoms, the FDA regulates them under the condom regulation, 884.5300. The same analysis applies to product codes MSC and OKW, which are also regulated in 884.5300.

**Figure 11 – Product Classification Pages for Product Codes HIS, NUC and QPD/Regulated in 884.5300**

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

New Search
Back to Search Results

|  |   |
|--|---|
| <b>Device</b>                          | Condom  |
| <b>Regulation Medical Specialty</b>    | Obstetrics/Gynecology   |
| <b>Review Panel</b>                    | Obstetrics/Gynecology   |
| <b>Product Code</b>                    | HIS   |
| <b>Premarket Review</b>                | <a href="#">Reproductive, Gynecology and Urology Devices (DHT3B)</a><br><a href="#">Reproductive, Gynecology and Urology Devices (DHT3B)</a>  |
| <b>Submission Type</b>                 | 510(k)  |
| <b>Regulation Number</b>               | 884.5300  |
| <b>Device Class</b>                    | 2   |
| <b>Total Product Life Cycle (TPLC)</b> | <a href="#">TPLC Product Code Report</a>  |
| <b>GMP Exempt?</b>                     | No  |
| <b>Summary Malfunction Reporting</b>   | Eligible  |
| <b>Implanted Device?</b>               | No  |
| <b>Life-Sustain/Support Device?</b>    | No  |
| <b>Recognized Consensus Standards</b>  | <ul style="list-style-type: none"> <li>● <a href="#">5-62 ASQ ANSI Z1.4-2003 (R2018) Sampling Procedures and Tables for Inspection by Attributes</a></li> <li>● <a href="#">6-214 ASTM D6355-07 (Reapproved 2017) Standard Test Method for Human Repeat Insult Patch Testing of Medical Gloves</a></li> <li>● <a href="#">6-411 ASTM D6499-18 Standard Test Method for Immunological Measurement of Antigenic Protein in Hevea Natural Rubber (HNR) and its Products</a></li> <li>● <a href="#">9-67 ASTM D7661-10 (Reapproved 2017) Standard Test Method for Determining Compatibility of Personal Lubricants with Natural Rubber Latex Condoms</a></li> <li>● <a href="#">9-86 ISO 16037 First edition 2002-05-15 (Amendment 1 2011-02-15) Rubber condoms for clinical trials - Measurement of physical properties [Including: Amendment (2011)]</a></li> <li>● <a href="#">9-111 ISO 4074 Third edition 2015-10-15 Natural latex rubber condoms - Requirements and test methods</a></li> <li>● <a href="#">9-115 ISO 29943-1 First edition 2017-07. Condoms -- Guidance on clinical studies -- Part 1: Male condoms, clinical function studies based on self-reports</a></li> <li>● <a href="#">9-117 ISO 16038 Second edition 2017-11 Male condoms - Guidance on the use of ISO 4074 and ISO 23409 in the quality management of condoms</a></li> <li>● <a href="#">9-120 ASTM D3492-16 Standard Specification for Rubber Contraceptives (Male Condoms)</a></li> </ul> |
| <b>Third Party Review</b>              | <ul style="list-style-type: none"> <li>● <a href="#">Eligible for 510(k) Third Party Review Program</a></li> </ul>  |
| <b>Accredited Persons</b>              | <ul style="list-style-type: none"> <li>● <a href="#">Center For Measurement Standards Of Industrial</a></li> <li>● <a href="#">Global Quality And Regulatory Services</a></li> <li>● <a href="#">Regulatory Technology Services, Llc</a></li> <li>● <a href="#">Third Party Review Group, Llc</a></li> </ul>  |

Page Last Updated: 05/22/2023

**Figure 11 – Product Classification Pages for Product Codes HIS, NUC and QPD/Regulated in 884.5300 – Continued**

### Product Classification

FDA Home Medical Devices Databases

New Search Back to Search Results

**Device**  
**Definition** Lubricant, Personal  
 This device is a personal lubricant, for penile and/or vaginal application, intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity, and supplement the body's natural lubrication. This product may or may not be compatible with natural rubber latex, polyisoprene, and/or polyurethane condoms.

**Regulation Medical Specialty** Obstetrics/Gynecology  
**Review Panel** Obstetrics/Gynecology  
**Product Code** NUC  
**Premarket Review** [Reproductive, Gynecology and Urology Devices \(DHT3B\)](#)  
 Reproductive, Gynecology and Urology Devices (DHT3B)

**Submission Type** 510(k)  
**Regulation Number** 884.5300  
**Device Class** 2  
**Total Product Life Cycle (TPLC)** [TPLC Product Code Report](#)  
**GMP Exempt?** No  
**Summary Malfunction Reporting** Eligible  
**Implanted Device?** No  
**Life-Sustain/Support Device?** No  
**Recognized Consensus Standard**  
 9-87 ASTM D7961-10 (Reapproved 2017)  
[Standard Test Method for Determining Compatibility of Personal Lubricants with Natural Rubber Latex Condoms](#)  
**Third Party Review**  
 Eligible for [510\(k\) Third Party Review Program](#)  
**Accredited Persons**  
 Center For Measurement Standards Of Industrial Global Quality And Regulatory Services  
 Regulatory Technology Services, Llc  
 Third Party Review Group, LLC

Page Last Updated: 05/22/2023

### Product Classification

FDA Home Medical Devices Databases

New Search Back to Search Results

**Device**  
**Definition** Personal Lubricant Ring  
 This device is a personal lubricant, for vaginal application, intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity, and supplement the body's natural lubrication. This product may or may not be compatible with natural rubber latex, polyisoprene, polyurethane, and/or internal condoms.

**Physical State** This device consists of a hollow plastic ring filled with a liquid vaginal lubricating solution.  
**Technical Method** The lubricating solution in the lumen of the ring is released through the wall of the tubing into the vagina, moisturizing and lubricating the vaginal mucosa.  
**Target Area** This device is intended to be used in the vagina.

**Regulation Medical Specialty** Obstetrics/Gynecology  
**Review Panel** Obstetrics/Gynecology  
**Product Code** QPD  
**Premarket Review** [Reproductive, Gynecology and Urology Devices \(DHT3B\)](#)  
 Reproductive, Gynecology and Urology Devices (DHT3B)

**Submission Type** 510(k)  
**Regulation Number** 884.5300  
**Device Class** 2  
**Total Product Life Cycle (TPLC)** [TPLC Product Code Report](#)  
**GMP Exempt?** No  
**Summary Malfunction Reporting** Ineligible  
**Implanted Device?** No  
**Life-Sustain/Support Device?** No  
**Third Party Review** Not Third Party Eligible

Page Last Updated: 05/22/2023

## Classification Example 2 – Devices with artificial intelligence

The figure below shows the major differences when using the Quick Search option and Advanced Search option using the words “artificial intelligence.”

**Figure 12 – Product Classification Database Comparison — Quick View (Top) vs. Advanced View (Bottom) Search Results for “Artificial Intelligence”**

### Product Classification

FDA Home Medical Devices Databases

1 to 9 of 9 Results for Artificial Intelligence Results per page 10

New Search Export To Excel Help

| Product Code        | Device   | Regulation Number                           | Device Class |
|---------------------|--|---|--------------|
| <a href="#">QJU</a> | <a href="#">Image Acquisition And/Or Optimization Guided By Artificial Intelligence</a>                    | 892.2100                                    | 2            |
| <a href="#">QAS</a> | <a href="#">Radiological Computer-Assisted Triage An ...</a>   | Radiological Computer Aided Triage And N... | 892.2080     |
| <a href="#">QFM</a> | <a href="#">Radiological Computer-Assisted Prioritiz ...</a>   | Radiological Computer Aided Triage And N... | 892.2080     |
| <a href="#">QVD</a> | <a href="#">Radiological Machine Learning Based Quantitative Imaging Software With Change Control Plan</a> | 892.2055                                    | 2            |
| <a href="#">QIH</a> | <a href="#">Automated Radiological Image Processing ...</a>  | Medical Image Management And Processing ... | 892.2050     |
| <a href="#">QKB</a> | <a href="#">Radiological Image Processing Software F ...</a>   | Medical Image Management And Processing ... | 892.2050     |
| <a href="#">PIB</a> | <a href="#">Diabetic Retinopathy Detection Device</a>  | Retinal Diagnostic Software Device          | 886.1100     |
| <a href="#">QNP</a> | <a href="#">Gastrointesinal Lesion Software Detectio ...</a>   | Gastrointestinal Lesion Software Detecti... | 876.1520     |
| <a href="#">QKQ</a> | <a href="#">Digital Pathology Image Viewing And Mana ...</a>   | Whole Slide Imaging System                  | 864.3700     |

Page Last Updated: 06/05/2023

Figure 12 – Product Classification Database Comparison — Quick View (Top) vs. Advanced View (Bottom) Search Results for “Artificial Intelligence” – Continued

## Product Classification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

| <a href="#">New Search</a>             | <a href="#">Back to Search Results</a>   |
|--|--|
| <b>Device Definition</b>               | Image Acquisition And/Or Optimization Guided By Artificial Intelligence<br><br>A radiological acquisition and/or optimization guidance system is a device that is intended to aid in the acquisition and/or optimization of images and/or diagnostic signals. The device interfaces with the acquisition system, analyzes its output, and provides guidance and/or feedback to the operator for improving image and/or signal quality. |
| <b>Physical State</b>                  | The subject software would utilize images acquired using an imaging system. The software can be installed on an existing imaging system, or can be operated on a computer that is connected to the imaging system.   |
| <b>Technical Method</b>                | The device's algorithm would be based on the analysis of images and/or diagnostic data. The underlying algorithms used for providing guidance to the users may be based on deep learning methods, trained on images obtained by trained operators.   |
| <b>Target Area</b>                     | Human body   |
| <b>Regulation Medical Specialty</b>    | Radiology  |
| <b>Review Panel</b>                    | Radiology  |
| <b>Product Code</b>                    | QJU  |
| <b>Premarket Review</b>                | <a href="#">Office of Radiological Health</a> (OHT8)<br>Division of Radiological Imaging and Radiation Therapy Devices (DHT8C)   |
| <b>Submission Type</b>                 | 510(k)   |
| <b>Regulation Number</b>               | 892.2100   |
| <b>Device Class</b>                    | 2  |
| <b>Total Product Life Cycle (TPLC)</b> | <a href="#">TPLC Product Code Report</a>   |
| <b>GMP Exempt?</b>                     | No   |
| <b>Summary Malfunction Reporting</b>   | Ineligible   |
| <b>Implanted Device?</b>               | No   |
| <b>Life-Sustain/Support Device?</b>    | No   |
| <b>Third Party Review</b>              | Not Third Party Eligible   |

Page Last Updated: 05/22/2023

In the example above, nine product codes were extracted from the database using the Quick Search option and just one product code (QJU) was extracted using the Advanced Search option. The product code QJU was also extracted using the Quick Search option. The regulation numbers were arranged in the first screenshot above to group the product codes by review panel. The first six product codes fall under Part 892 Radiology, covering the following four regulation numbers:

- 892.2100 – The regulation name and description are pending with the FDA as product code QJU (image acquisition and/or optimization guided by artificial intelligence), and the regulation number were established under a De Novo pathway (DEN190040).<sup>3</sup> There are no other product codes in this regulation.
- 892.2080 – Radiological computer-aided triage and notification software: Product code QAS (radiological computer-assisted triage and notification software) and the regulation number were established under a De Novo pathway (DEN170073).<sup>4</sup> Product code QFM (radiological computer-assisted prioritization software for lesions) was added to the regulation with a different name and technical method. There are no other product codes in this regulation.
- 892.2055 – The regulation name and description are pending with the FDA as product code QVD (radiological machine learning-based quantitative imaging software with change control plan), and the regulation number were established under a De Novo pathway (DEN220063).<sup>5</sup> There are no other product codes in this regulation.

- 892.2050** – Medical image management and processing system: Product codes QIH (automated radiological image processing software) and QKB (radiological image processing software for radiation therapy) were added to the regulation, each with a different device name than what appears in the regulation and with added definitions and technical methods on their respective product classification pages that do not appear in the original regulation. There are 11 product codes in this regulation, as seen in **Figure 13**. Artificial intelligence is not referenced in the regulation or in the names for product codes QIH and QKB. However, artificial intelligence is mentioned in the definition and technical method for both of these product codes on their respective product classification pages. The intended use of these two devices is reflected in their respective definitions.

Figure 13 – Product Codes in 892.2050

**Product Classification**  
 FDA Home Medical Devices Databases

1 to 11 of 11 results  
 892.2050

Results per page 25

[New Search](#) [Export to Excel](#) [Help](#)

| Product Code | Device   | Regulation Number        | Device Class |
|--------------|--|--------------------------|--------------|
| PZO          | <a href="#">Software For Visualization Of Vascular Anatomy And ...</a> | <a href="#">892.2050</a> | 2            |
| QQE          | <a href="#">Image Management Software For Planning Of Otologic ...</a> | <a href="#">892.2050</a> | 2            |
| NFJ          | <a href="#">System, Image Management, Ophthalmic</a>                   | <a href="#">892.2050</a> | 2            |
| LLZ          | <a href="#">System, Image Processing, Radiological</a>                 | <a href="#">892.2050</a> | 2            |
| NWE          | <a href="#">Colon Computed Tomography System, Computer Aided D ...</a> | <a href="#">892.2050</a> | 2            |
| OEB          | <a href="#">Lung Computed Tomography System, Computer-Aided De ...</a> | <a href="#">892.2050</a> | 2            |
| OMJ          | <a href="#">Chest X-Ray Computer Aided Detection</a>                   | <a href="#">892.2050</a> | 2            |
| PGY          | <a href="#">Display, Diagnostic Radiology</a>                          | <a href="#">892.2050</a> | 2            |
| QIH          | <a href="#">Automated Radiological Image Processing Software</a>       | <a href="#">892.2050</a> | 2            |
| QKB          | <a href="#">Radiological Image Processing Software For Radiati ...</a> | <a href="#">892.2050</a> | 2            |
| QTZ          | <a href="#">Radiological Image Processing Software For Ablatio ...</a> | <a href="#">892.2050</a> | 2            |

Page Last Updated: 05/22/2023

The last three product codes in the table from **Figure 12** fall under Part 886 Ophthalmic, Part 876 Gastroenterology/Urology and Part 864 Pathology, covering the following three regulations:

- **886.1100** – Retinal diagnostic software device: Product code PIB (diabetic retinopathy detection device) and the regulation were established under a De Novo pathway (DEN 180001).<sup>6</sup> Artificial intelligence is referenced in the technical method for this device on its product classification page. The intended use of this device is reflected in its definition. There are no other product codes in this regulation.
- **876.1520** – Gastrointestinal lesion software detection system: Product code QNP (gastrointestinal lesion software detection system) and the regulation were established under a De Novo pathway (DEN200055).<sup>7</sup> Artificial intelligence is referenced in the technical method for this device on its product classification page. The intended use of this device is reflected in its definition. There are no other product codes in this regulation.
- **864.3700** – Whole slide imaging system: Product code QKQ (digital pathology image viewing and management software) was added to the regulation with a different device name, definition, physical state, technical method and target area than what appears for product code PSY (whole slide imaging system) that established this regulation under a De Novo pathway (DEN160056).<sup>8</sup> Artificial intelligence is referenced in the technical method for product code QKQ on its product classification page. The intended use of this device is reflected in its definition. There is one other product code, PZZ (digital pathology display), in this regulation.

From Example 2, we see that there are several devices that all have some form of artificial intelligence with different intended uses and, as such, have been classified by the FDA under nine different products codes covering seven different regulation numbers and four review panels (Radiology, Ophthalmic, Gastroenterology/Urology and Pathology) as Class II devices; five of the product codes and regulation numbers have been established under De Novo pathways.

It should be noted that only the product code and regulation are searched, and not the individual devices; not all product codes and regulations reference a specific technology. Therefore, while this is a good beginning point for researching devices with artificial intelligence, it should not be the only search, and there may be other devices with artificial intelligence that were not found in the search.

### Classification Example 3 – Orthopedic implants — shoulder prosthesis

Few medical devices at the Center for Device and Radiological Health (CDRH) are more complicated to correctly classify than prosthetic hip, knee and shoulder systems because of their different intended uses, indications for use, design features and the components/materials used to make them. These complications can result in some device systems requiring PMA submissions while other device systems require 510(k) submissions. For this classification example, prosthetic shoulder systems were selected because there are fewer product codes to describe these device systems than there are for prosthetic hip and knee systems.

The figure below shows the subtle differences when using the Quick Search option and Advanced Search option using the words “shoulder prosthesis.”

Figure 14 – Product Classification Database Comparison — Quick View (Top) vs. Advanced View (Bottom) Search Results for “Shoulder Prosthesis”

**Product Classification**

FDA Home Medical Devices Databases

1 to 11 of 11 Results for Shoulder Prosthesis Results per page 25

New Search Export To Excel Help

| Product Code        | Device  | Regulation Number | Device Class |
|---------------------|---|-------------------|--------------|
| <a href="#">KWR</a> | <a href="#">Prosthesis, Shoulder, Constrained, Metal...</a> Shoulder Joint Metal/Metal Or Metal/Poly...             | 888.3640          | 3            |
| <a href="#">KYM</a> | <a href="#">Metallic Cemented Glenoid Hemi-Shoulder ...</a> Shoulder Joint Glenoid (Hemi-Shoulder) M...             | 888.3680          | 3            |
| <a href="#">MJT</a> | <a href="#">Prosthesis, Shoulder, Humeral (Bipolar Hemi-Shoulder) Metal/Polymer, Cemented Or Uncemented</a>         |                   | 3            |
| <a href="#">QKW</a> | <a href="#">Shoulder Joint Humeral (Hemi-Shoulder) Ceramic Head/Metallic Stem Cemented Or Uncemented Prosthesis</a> | 888.3695          | 2            |
| <a href="#">HSD</a> | <a href="#">Prosthesis, Shoulder, Hemi- Humeral, Me ...</a> Shoulder Joint Humeral (Hemi-Shoulder) M...             | 888.3690          | 2            |
| <a href="#">KWT</a> | <a href="#">Prosthesis, Shoulder, Non-Constrained, M ...</a> Shoulder Joint Metal/Polymer Non-Constra...            | 888.3650          | 2            |
| <a href="#">KWS</a> | <a href="#">Prosthesis, Shoulder, Semi-Constrained, ...</a> Shoulder Joint Metal/Polymer Semi-Constr...             | 888.3660          | 2            |
| <a href="#">PAO</a> | <a href="#">Prosthesis, Shoulder, Semi-Constrained, ...</a> Shoulder Joint Metal/Polymer Semi-Constr...             | 888.3660          | 2            |
| <a href="#">PHX</a> | <a href="#">Shoulder Prosthesis, Reverse Configurati ...</a> Shoulder Joint Metal/Polymer Semi-Constr...            | 888.3660          | 2            |
| <a href="#">PKC</a> | <a href="#">Prosthesis, Total Anatomic Shoulder, Unc ...</a> Shoulder Joint Metal/Polymer Semi-Constr...            | 888.3660          | 2            |
| <a href="#">MBF</a> | <a href="#">Prosthesis, Shoulder, Semi-Constrained, ...</a> Shoulder Joint Metal/Polymer/Metal Nonco...             | 888.3670          | 2            |

Page Last Updated: 05/22/2023

**Product Classification**

FDA Home Medical Devices Databases

1 to 14 of 14 results for Shoulder Prosthesis Results per page 25

New Search Export To Excel Help

| Product Code        | Device  | Regulation Number        | Device Class |
|---------------------|---|--------------------------|--------------|
| <a href="#">KWR</a> | <a href="#">Prosthesis, Shoulder, Constrained, Metal/Metal Or ...</a> Shoulder Joint Metal/Metal Or Metal/Poly...   | <a href="#">888.3640</a> | 3            |
| <a href="#">KYM</a> | <a href="#">Metallic Cemented Glenoid Hemi-Shoulder Prosthesis</a> Shoulder Joint Glenoid (Hemi-Shoulder) M...      | <a href="#">888.3680</a> | 3            |
| <a href="#">MJT</a> | <a href="#">Prosthesis, Shoulder, Humeral (Bipolar Hemi-Shoulder) Metal/Polymer, Cemented Or Uncemented</a>         |                          | 3            |
| <a href="#">HSD</a> | <a href="#">Prosthesis, Shoulder, Hemi- Humeral, Metallic Unc ...</a> Shoulder Joint Humeral (Hemi-Shoulder) M...   | <a href="#">888.3690</a> | 2            |
| <a href="#">KWS</a> | <a href="#">Prosthesis, Shoulder, Semi-Constrained, Metal/Poly ...</a> Shoulder Joint Metal/Polymer Semi-Constr...  | <a href="#">888.3660</a> | 2            |
| <a href="#">KWT</a> | <a href="#">Prosthesis, Shoulder, Non-Constrained, Metal/Polym ...</a> Shoulder Joint Metal/Polymer Non-Constra...  | <a href="#">888.3650</a> | 2            |
| <a href="#">MBF</a> | <a href="#">Prosthesis, Shoulder, Semi-Constrained, Metal/Polym ...</a> Shoulder Joint Metal/Polymer/Metal Nonco... | <a href="#">888.3670</a> | 2            |
| <a href="#">PAO</a> | <a href="#">Prosthesis, Shoulder, Semi-Constrained, Metal/Polym ...</a> Shoulder Joint Metal/Polymer Semi-Constr... | <a href="#">888.3660</a> | 2            |
| <a href="#">PHX</a> | <a href="#">Shoulder Prosthesis, Reverse Configuration</a> Shoulder Joint Metal/Polymer Semi-Constr...              | <a href="#">888.3660</a> | 2            |
| <a href="#">PKC</a> | <a href="#">Prosthesis, Total Anatomic Shoulder, Uncemented Me ...</a> Shoulder Joint Metal/Polymer Semi-Constr...  | <a href="#">888.3660</a> | 2            |
| <a href="#">QHE</a> | <a href="#">Shoulder Arthroplasty Implantation System</a> Shoulder Joint Metal/Polymer Semi-Constr...               | <a href="#">888.3660</a> | 2            |
| <a href="#">QHQ</a> | <a href="#">Total Shoulder Arthroplasty System</a> Shoulder Joint Metal/Polymer Semi-Constr...                      | <a href="#">888.3660</a> | 2            |
| <a href="#">QKW</a> | <a href="#">Shoulder Joint Humeral (Hemi-Shoulder) Ceramic Head/Metallic Stem Cemented Or Uncemented Prosthesis</a> | 888.3695                 | 2            |
| <a href="#">PAE</a> | <a href="#">Upper Extremity Prosthesis With Multiple Simultane ...</a> Upper Extremity Prosthesis Including A S...  | <a href="#">890.3450</a> | 2            |

Page Last Updated: 05/22/2023

In the example above, we see that 11 product codes were extracted from the database using the Quick Search option, and 14 product codes were extracted using the Advanced Search option. Product codes QHE, QHQ and PAE were not extracted using the Quick Search option. In looking further at all the product codes that were extracted using the advanced search option in the second screenshot above, only product code PAE falls under Part 890 Physical Medicine — not Part 888 Orthopedic, the panel that reviews implantable orthopedic devices.

The other 13 product codes shown in **Figure 14** in this search example cover the following seven regulation numbers, while product code MJT (prosthesis, shoulder, humeral (bipolar hemi-shoulder) metal/polymer, cemented or uncemented) is a Class III device that has not been assigned a regulation number, and its current submission type described on its product classification page requires communication with the Office of Device Evaluation to confirm its submission type.

- **888.3640** – Shoulder joint metal/metal or metal/polymer constrained cemented prosthesis: Product code KWR (prosthesis, shoulder, constrained, metal/metal or metal/polymer cemented) has the same name as the regulation and is a Class III device that requires FDA approval under a PMA submission. There are no other product codes in this regulation.
- **888.3650** – Shoulder joint metal/polymer non-constrained cemented prosthesis: Product code KWT (prosthesis, shoulder, non-constrained, metal/polymer cemented) has the same name as the regulation and is a Class II device that requires FDA clearance under a 510(k) submission. There are no other product codes in this regulation.
- **888.3660** – Shoulder joint metal/polymer semi-constrained cemented prosthesis: Product code KWS (prosthesis, shoulder, non-constrained, metal/polymer cemented) has a different name than the regulation and is a Class II device that requires FDA clearance under a 510(k) submission. There are five other product codes in this regulation — PAO (prosthesis, shoulder, semi-constrained, metal/polymer with additive, cemented), PHX (prosthesis, shoulder, semi-constrained, metal/polymer with additive, cemented), PKC (prosthesis, total anatomic shoulder, uncemented metaphyseal humeral stem with no diaphyseal incursion, semi-constrained), QHE (shoulder arthroplasty implantation system) and QHQ (total shoulder arthroplasty system) — and each has a different name than the regulation name. They also have additional descriptive information (definition, physical state and technical method) on their respective product classification pages.
- **888.3670** – Shoulder joint metal/polymer/metal non-constrained or semi-constrained porous-coated uncemented prosthesis: Product code MBF (prosthesis, shoulder, semi-constrained, metal/polymer, uncemented) has a similar name as the regulation and is a Class II device that requires FDA clearance under a 510(k) submission. There are no other product codes in this regulation.
- **888.3680** – Shoulder joint glenoid (hemi-shoulder) metallic cemented prosthesis: Product code KYM (metallic cemented glenoid hemi-shoulder prosthesis) has the same name as the regulation and is a Class III device that requires FDA approval under a PMA submission. There are no other product codes in this regulation.
- **888.3690** – Shoulder joint humeral (hemi-shoulder) metallic uncemented prosthesis: Product code HSD (prosthesis, shoulder, hemi-, humeral, metallic uncemented) has the same name as the regulation and is a Class II device that requires FDA clearance under a 510(k) submission. There are no other product codes in this regulation.
- 888.3695 – The regulation name and description are pending as product code QKW (shoulder joint humeral (hemi-shoulder) ceramic head/metallic stem cemented or uncemented prosthesis) were established under a De Novo pathway (DEN220012).<sup>9</sup> There are no other product codes in this regulation.

From Example 3, we see that there are 13 product codes for the distinct types of prosthetic shoulder systems covering seven different regulation numbers that are all reviewed by the Orthopedic device panel as either Class II or Class III devices subject to 510(k) and PMA submissions, respectively. When trying to correctly classify prosthetic shoulder systems, as well as hip and knee systems, it is crucial to fully understand the overall design of these complicated joint systems as well as the main components and materials used to make them. Therefore, the different regulation number descriptions as well as the product classification pages need to be reviewed and understood, especially when the product classification page has additional descriptive information related to the device name, definition, physical state, technical method and target area (which some product codes do not have), as this additional information is not always found in regulation number descriptions.

# Using other FDA databases

Several other FDA databases within the CDRH that were identified above can also be used to classify or further confirm the classification of your device. These databases are best used if you know of another device that is the same as or like your device in terms of its intended use and technology and that has been listed, cleared and/or approved by the FDA. In this section, we will focus on the FDA's Registration and Listing (R&L) database and their 510(k) database. Using the FDA's De Novo, PMA and HDE databases is similar to using the 510(k) database.





## Classification Example 4 – Using the FDA R&L database to confirm classification

The FDA’s R&L database, as seen in **Figure 15**, is best used when you have a high degree of confidence that your device is the same as or like another legally marketed device in the U.S. with respect to its intended use and technology. It is the manufacturer’s responsibility to ensure their device has been correctly classified and listed with the FDA, as seen below.

Figure 15 – FDA Establishment R&L Database

### Establishment Registration & Device Listing

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

This database includes:

- medical device manufacturers registered with FDA and
- medical devices listed with FDA

**Note:** Registration of a device establishment, assignment of a registration number, or listing of a medical device does not in any way denote approval of the establishment or its products by FDA.

[Learn More...](#)

**Search Database**

[Help](#)
[Download Files](#)

|  |   |
|--|---|
| Establishment or Trade Name <input style="width: 90%;" type="text"/> | Registration or FEI Number <input style="width: 90%;" type="text"/> |
| Owner/Operator Name <input style="width: 90%;" type="text"/>         | Owner/Operator Number <input style="width: 90%;" type="text"/>      |
| Proprietary Name <input style="width: 90%;" type="text"/>            | Classification Device Name <input style="width: 90%;" type="text"/> |
| Product Code <input style="width: 20%;" type="text"/>                | Establishment Type <input style="width: 90%;" type="text"/>         |
| Establishment State (U.S.) <input style="width: 90%;" type="text"/>  | Establishment Country * <input style="width: 90%;" type="text"/>    |

[Quick Search](#)

[Clear Form](#)

**Need to update your information?** To modify, add, or delete information, [log onto your FURLS account](#).

*The changes will appear in the public registration and listing database when it is updated. Please note: There is a delay between the time that the data is uploaded and the time that the data appears in the public registration and listing database. Existing device listings may also be effected by the update and may not be fully viewable until after the update has been completed. Updates are generally completed weekly by Monday evening.*

\* Category includes Countries, States and Regions

Page Last Updated: 05/22/2023

## WHITE PAPER

In this example, a new device company wants to manufacture and market hand-held electric massagers in the U.S. and sell them under their own brand name for relief of minor muscle aches and pains to a targeted area of the body. The company has identified product code ISA (massager, therapeutic, electric) in the product classification database as the likely product code, which is a Class I 510(k)-exempt device under [890.5660](#). To confirm the device classification, the company conducts a search using the FDA's R&L database under product code ISA, which lists 100 different establishment types — e.g., manufacturers, contract manufacturers, specification developers, etc. — registered with the FDA selling electric massagers listed under product code ISA.

In looking closer at the R&L pages for several different establishments, the new company noticed that two companies show their therapeutic electric massagers as also having been cleared by the FDA under two 510(k)s ([K210166](#) and [K173692](#)), as seen in **Figure 16**. Upon further review of the two 510(k) summary documents in the 510(k) database, they found that these two electric massagers had broader intended use claims beyond relief of minor muscle aches and pains, including temporary increase in local blood circulation and activation of connective tissue, as well as incorporating more advanced technology than what is typically found in a hand-held electric massager.

Specifically, the device cleared under K210166 uses an electro-hydraulic method of creating shock waves to achieve its intended principles of operation, while the device cleared under K173692 uses pressurized air pulses to achieve its intended principles of operation. These technical characteristics are different than traditional electromechanically induced vibrations produced by the new company's device. Based on this review, the new company concluded that a 510(k) did not need to be submitted to the FDA for their device as the limitations to 510(k) exemption as described in parts (a) and (b) of 890.9 did not apply to the intended use and technology of their therapeutic electric massager.



Figure 16 – FDA Establishment R&L Pages for Two ISA Devices (K210166 and K173692)

## Establishment Registration & Device Listing

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[New Search](#) [Back To Search Results](#)

|   |                                 |
|---|---------------------------------|
| <b>Proprietary Name:</b>                | Omnispec™ ED1000                |
| <b>Classification Name:</b>             | MASSAGER, THERAPEUTIC, ELECTRIC |
| <b>Product Code:</b>                    | <a href="#">ISA</a>             |
| <b>Device Class:</b>                    | 1                               |
| <b>Regulation Number:</b>               | <a href="#">890.5660</a>        |
| <b>Medical Specialty:</b>               | Physical Medicine               |
| <b>Registered Establishment Name:</b>   | <a href="#">MEDISPEC, LTD.</a>  |
| <b>Registered Establishment Number:</b> | 3002807616                      |
| <b>Premarket Submission Number:</b>     | <a href="#">K210166</a>         |
| <b>Owner/Operator:</b>                  | <a href="#">MEDISPEC, LTD.</a>  |
| <b>Owner/Operator Number:</b>           | 9006433                         |
| <b>Establishment Operations:</b>        | Manufacturer                    |

Page Last Updated: 05/22/2023

---

## Establishment Registration & Device Listing

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[New Search](#) [Back To Search Results](#)

|   |  |
|---|--|
| <b>Proprietary Name:</b>                | CHATTANOOGA Mobile 2 RPW USA; D-ACTOR 100; D-ACTOR 200; D-ACTOR 50; DUOLITH SD1 R-SW; MASTERPULS ONE |
| <b>Classification Name:</b>             | MASSAGER, THERAPEUTIC, ELECTRIC  |
| <b>Product Code:</b>                    | <a href="#">ISA</a>  |
| <b>Device Class:</b>                    | 1  |
| <b>Regulation Number:</b>               | <a href="#">890.5660</a>   |
| <b>Medical Specialty:</b>               | Physical Medicine  |
| <b>Registered Establishment Name:</b>   | <a href="#">STORZ MEDICAL AG</a>   |
| <b>Registered Establishment Number:</b> | 9613347  |
| <b>Premarket Submission Number:</b>     | <a href="#">K173692</a>  |
| <b>Owner/Operator:</b>                  | <a href="#">STORZ MEDICAL AG</a>   |
| <b>Owner/Operator Number:</b>           | 9028780  |
| <b>Establishment Operations:</b>        | Manufacturer   |

Page Last Updated: 05/22/2023



## Classification Example 5 – Using the FDA 510(k) database

The FDA's 510(k) database is best used when you have a high degree of confidence that your device is the same as or like another legally marketed device in the U.S. with respect to its intended use and technology. However, in this example, there are some differences in the technology used in the new company's device when compared to most predicate devices under the same product code, so the manufacturer wants to see if they can narrow down their search effort to confirm whether they can still submit a 510(k) because of the difference in technology in their device, as opposed to having to submit a De Novo.

In this example, the manufacturer produces dental abutments (product code NHA) that are used with dental implants (product code DZE). However, instead of making their dental abutments out of titanium or ceramic, which are the materials used to make most dental abutments, they make them out of the polymer, polyetheretherketone (PEEK).

Knowing that there are more than 500 dental abutments cleared by the FDA, as shown in the 510(k) database when searching using product code NHA, the manufacturer did not want to waste a lot of time looking through dozens of 510(k) summary documents to try to locate a dental abutment made of PEEK that could be used as a predicate device in a future 510(k) submission.

Setting up the predicate device search, the manufacturer used product code NHA, entered the word "PEEK" in the device name field and selected the word "Dental" from the drop-down menu for the device review panel, as seen in **Figure 17**. This search produced five possible predicate devices, seen in **Figure 18**, where the word "PEEK" was found in the trade of the device that was submitted to the FDA. It is possible that other dental abutments have been cleared that are made from PEEK, but if "PEEK" does not appear in the trade name of the device, it is harder to locate such a predicate device without having to look through many 510(k) summary documents.

Figure 17 – FDA 510(k) Database

## 510(k) Premarket Notification

FDA Home Medical Devices Databases

A 510(K) is a premarket submission made to FDA to demonstrate that the device to be marketed is as safe and effective, that is, substantially equivalent, to a legally marketed device (section 513(i)(1)(A) FD&C Act) that is not subject to premarket approval.

[Learn more...](#)

### Search Database

Help Download Files

510K Number  Type   Product Code

Center   Combination Products

Applicant Name  Cleared/Approved In Vitro Products

Device Name  Redacted FOIA 510(k)

Panel   Third Party Reviewed

Decision

Decision Date  to  Clinical Trials

Sort by

[Quick Search](#) [Clear Form](#)

Figure 18 – FDA 510(k) Database Search Results Using “NHA,” “PEEK” and “Dental”

## 510(k) Premarket Notification

FDA Home Medical Devices Databases

1 to 5 of 5 Results  
 Panel: Dental ProductCode: NHA Device Name: PEEK Decision Date To: 05/22/2023

Results per Page

[New Search](#)  Export to Excel | [Download Files](#) | [More About 510\(k\)](#)

| Device Name  | Applicant   | 510(K) Number           | Decision Date |
|--|---|-------------------------|---------------|
| <a href="#">Blue Sky Bio Multi One Implant System, Blue Sky Bio Long Implant System, Blue Sky Bio Peek Temporary Abutments</a> | Blue Sky Bio, LLC                                   | <a href="#">K212785</a> | 06/30/2022    |
| <a href="#">Southern Implants Peek Abutments</a>   | Southern Implants (Pty) Ltd                         | <a href="#">K191250</a> | 12/05/2019    |
| <a href="#">Southern Implants Peek Abutments</a>   | Southern Implants (Pty) Ltd                         | <a href="#">K172160</a> | 02/09/2018    |
| <a href="#">Neodent Implant System - Cm Pro Peek Abutment</a>  | JJGC Industria E Comercio De Materiais Dentarios SA | <a href="#">K170080</a> | 11/02/2017    |
| <a href="#">Nobelprocera Peek Abutments</a>  | NOBEL BIO CARE AB                                   | <a href="#">K120954</a> | 06/05/2013    |

Page Last Updated: 05/22/2023

# Some helpful tips for classifying your device

## Locating all product codes in a regulation number

As we have seen in some of the examples above, there can often be 10 or more product codes in the same regulation, and you want to make sure you have selected the correct product code for your device. If you are not sure about the product code of your medical device but are certain of the regulation number that your device falls under based on its intended use and technological characteristics, and

assuming that a regulation number has been established by the FDA for your device, insert the regulation number in the regulation number field on the product classification page and all of the product codes in that regulation will be extracted, as seen in **Figure 19** using [CFR 880.5725 Infusion Pump](#). In this example, 17 product codes are extracted for “infusion pump,” as seen in **Figure 20**.

Figure 19 – Product Classification Database Using Regulation Number 880.5725

**Product Classification**  
FDA Home Medical Devices Databases

This database includes:

- a list of all medical devices with their associated classifications, product codes, FDA Premarket Review organizations, and other regulatory information.

[learn more...](#)

**Search Database** Help Download Files

Device  Product Code

Review Panel  Regulation Number

Submission Type  Third Party Eligible

Implanted Device  Life-Sustain/Support Device  Device Class

Summary Malfunction Reporting

[Go to Quick Search](#) [Clear Form](#)

Figure 20 – Product Classification Database Results Using Regulation Number 880.5725

**Product Classification**  
 FDA Home Medical Devices Databases  
 1 to 17 of 17 results  
 880.5725 Results per page 25

[New Search](#) [Export to Excel](#) [Help](#)

| Product Code | Device  | Regulation Number | Device Class |
|--------------|---|-------------------|--------------|
| LZF          | <a href="#">Pump, Infusion, Analytical Sampling</a>               | 880.5725          | 2            |
| LZG          | <a href="#">Pump, Infusion, Insulin</a>                           | 880.5725          | 2            |
| FRN          | <a href="#">Pump, Infusion</a>                                    | 880.5725          | 2            |
| LDR          | <a href="#">Controller, Infusion, Intravascular, Electronic</a>   | 880.5725          | 2            |
| LGZ          | <a href="#">Warmer, Thermal, Infusion Fluid</a>                   | 880.5725          | 2            |
| LHF          | <a href="#">Warmer, Microwave, Infusion Fluid</a>                 | 880.5725          | 2            |
| LZH          | <a href="#">Pump, Infusion, Enteral</a>                           | 880.5725          | 2            |
| MEA          | <a href="#">Pump, Infusion, Pca</a>                               | 880.5725          | 2            |
| MEB          | <a href="#">Pump, Infusion, Elastomeric</a>                       | 880.5725          | 2            |
| MHD          | <a href="#">Pump, Infusion, Gallstone Dissolution</a>             | 880.5725          | 2            |
| MRZ          | <a href="#">Accessories, Pump, Infusion</a>                       | 880.5725          | 2            |
| OPP          | <a href="#">Pump, Infusion, Insulin Bolus</a>                     | 880.5725          | 2            |
| PHC          | <a href="#">Infusion Safety Management Software</a>               | 880.5725          | 2            |
| PKP          | <a href="#">Immunoglobulin G (Igg) Infusion System</a>            | 880.5725          | 2            |
| PMS          | <a href="#">Peripheral Intravenous (Piv) Infiltration Monitor</a> | 880.5725          | 2            |
| QJY          | <a href="#">Infusion Pump, Drug Specific, Pharmacy-Filled</a>     | 880.5725          | 2            |
| MRH          | <a href="#">Pump, Infusion, Ophthalmic</a>                        | 880.5725          | 2            |

Page Last Updated: 05/22/2023

In this example, unless a manufacturer is making the most common type of bedside infusion pump — typically found in a hospital or similar healthcare facility — that is classified under product code FRN (pump, infusion), the manufacturer will need to look more closely at the other, more specialized types of infusion pumps that fall under product codes LZG, LZH, MEA, MEB, OPP, PKP, QJY and MRH to determine whether their device is more accurately described under one of these other product codes. In this case, which will likely require looking at 510(k) summary documents to read the description of those devices that are

classified under product codes LZG, LZH, MEA, MEB and MRH, as there is no additional information in the form of definition, physical state, technical method and target area for these devices on their respective product classification pages like there is for product codes OPP, PKP and QJY. Part (a) of 880.5725 describes the most common type of infusion pump, product code FRN, while product codes LZG, LZH, MEA, MEB, MRH, OPP, PKP, QJY, and MRH are other types of devices or accessories that are used with infusion pumps as identified by their product code names, seen in the table in **Figure 20**.

## Additional information on the product classification page

As previously discussed in this white paper, when a product classification page has additional information for definition, physical state, technical method and target area for a particular product code, it likely means one of the following:

- The product code and regulation number were established under a De Novo pathway. This can easily be confirmed by placing the product code for the device in the product code field in the [De Novo database](#) and running a search to see if a DEN number is found, as seen in Figure 21 for product code [QMJ](#) (powered radiofrequency toothbrush) and CFR 872.6866.
- The FDA added the product code to an existing regulation number that has more than one product code in that regulation.
- The FDA has not written a regulation for the product code that applies to:
  - Unclassified (pre-amendment) devices
  - Not-classified devices
  - HDE devices
  - Certain Class III devices

Figure 21 – De Novo Database Results Using Product Code QMJ

### Device Classification Under Section 513(f)(2)(De Novo)

[FDA Home](#) [Medical Devices](#) [Databases](#)

1 result found  
productcode: [QMJ](#) Decision Date To: 05/22/2023

results per page


| <a href="#">New Search</a> |                        | <a href="#">Download Files</a>   <a href="#">More About De Novo</a> |               |               |
|----------------------------|------------------------|---|---------------|---------------|
| Device Name                | Requester              | De Novo Number  | 510(k) Number | Decision Date |
| <a href="#">ToothWave™</a> | Home Skinovations Ltd. | <a href="#">DEN190039</a>   |               | 09/17/2020    |

Page Last Updated: 05/22/2023

---

### Device Classification Under Section 513(f)(2)(De Novo)

[FDA Home](#) [Medical Devices](#) [Databases](#)



[510\(k\)](#) | [DeNovo](#) | [Registration & Listing](#) | [Adverse Events](#) | [Recalls](#) | [PMA](#) | [HDE](#) | [Classification](#) | [Standards](#)  
[CFR Title 21](#) | [Radiation-Emitting Products](#) | [X-Ray Assembler](#) | [Medsun Reports](#) | [CLIA](#) | [TPLC](#)

| <a href="#">New Search</a>  | <a href="#">Back to Search Results</a>   |
|---|--|
| <p><b>Device Classification Name</b></p> <p><b>De Novo Number</b></p> <p><b>Device Name</b></p> <p><b>Requester</b></p> <p><b>Contact</b></p> <p><b>Regulation Number</b></p> <p><b>Classification Product Code</b></p> <p><b>Date Received</b></p> <p><b>Decision Date</b></p> <p><b>Decision</b></p> <p><b>Classification Advisory Committee</b></p> <p><b>Review Advisory Committee</b></p> <p><b>Reclassification Order</b></p> <p><b>FDA Review Type</b></p> | <p><a href="#">Powered Radiofrequency Toothbrush</a></p> <p>DEN190039</p> <p>ToothWave™</p> <p>Home Skinovations Ltd.<br/>Tabor Building, Shaar Yokneam<br/>Yokneam Illit, IL 2069200</p> <p>Amit Goren</p> <p><a href="#">872.6866</a></p> <p><a href="#">QMJ</a></p> <p>08/22/2019</p> <p>09/17/2020</p> <p>Granted (DENG)</p> <p>Dental</p> <p>Dental</p> <p><a href="#">Reclassification Order</a></p> <p><a href="#">Decision Summary</a></p> <p>Direct</p> |



## Introducing new technology to an existing class of device

When innovative technology is introduced to an existing class of device that is low-to moderate-risk, there is the potential for a De Novo pathway to be required by the FDA to commercialize that device in the U.S. Product code QMJ (powered radiofrequency toothbrush) and CFR 872.6866 as discussed above and seen in **Figure 21** is a good example of this.

In the above example for the powered radiofrequency toothbrush, the device manufacturer added radiofrequency technology to a powered toothbrush. A powered toothbrush is a Class I 510(k)-exempt device under product code JEQ (toothbrush, powered) and regulation number [872.6865](#), but subject to the 510(k) limitations in CFR 872.9 for dental devices. The intended use of a toothbrush (manual, powered or powered with radiofrequency technology) is to remove adherent plaque and food debris from the teeth to reduce tooth decay. While it's possible the FDA could have requested that the manufacturer of the powered radiofrequency toothbrush submit a 510(k) for their device under product code JEQ because its technology operates using a different fundamental scientific technology than a legally marketed powered toothbrush, it appears that the radiofrequency technology used in this device was different enough to require a De Novo be submitted to the FDA, thus creating the new Class II device under product code QMJ and CFR 872.6866.



# Summary and conclusion



The FDA's medical device classification system is one of the most complicated when compared to other countries, with 16 review panels, more than 6,500 product codes, more than 1,700 regulation numbers, and devices identified as unclassified, Class I, Class II, Class III, not classified and HDE, with eight routes to commercialization in the U.S. – 510(k)-exempt, 510(k), De Novo, PMA, HDE, EUA, enforcement discretion and contact ODE. Understanding how your device is classified is the first step toward ensuring that you have identified the correct route for commercializing it in the U.S. and have developed a successful regulatory strategy.

While many devices are easy to classify in the FDA's device classification system, many are not. All device classifications need to be carefully evaluated before proceeding with a regulatory submission to the FDA. Emergo by UL can assist you in classifying your medical device or contacting the FDA for assistance under their **513(g) Request for Information**<sup>10</sup> (classification) process.

## End Notes

1. At the time this White Paper was published by Emergo by UL (May 2023).
2. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/cfrsearch.cfm?fr=884.5305>
3. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/cfrsearch.cfm?fr=884.5305>
4. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/denovo.cfm?id=DEN170073>
5. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/cfrsearch.cfm?fr=892.2055>
6. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/denovo.cfm?id=DEN180001>
7. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/denovo.cfm?id=DEN200055>
8. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/denovo.cfm?id=DEN160056>
9. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/cfrsearch.cfm?fr=888.3695>
10. <https://www.fda.gov/media/78456/download>

# About the author

**Stuart R. Goldman** has over thirty years of combined RA/QA experience in medical devices, including fifteen years in industry working on high-risk Class III/II implantable devices. At Emergo by UL, Stuart focuses on the United States market and has extensive expertise in device classification and testing requirements; regulatory pathway strategies; Q-Submissions and clearance of over forty 510(k)s and submission of over twenty 513(g)s.



**[EmergobyUL.com](https://www.emergobyul.com)**

Emergo by UL and the Emergo by UL logo are trademarks of Emergo Global Consulting LLC © 2023. All rights reserved. This document may not be copied or distributed, in whole or in part, without express written permission from Emergo by UL. It is provided for general information purposes only and is not intended to convey legal or other professional advice.